



CO-ORDINATION OF NOTIFIED BODIES
PPE Regulation 2016/425

PPE-R/05.26-013
Version 02

RECOMMENDATION FOR USE

| | | |
|---|--|--------------|
| Number of pages: 1 | Approval stage: | Approved on: |
| Origin: Vertical Group 5 | <input checked="" type="checkbox"/> Vertical Group | 16.06.2021 |
| | <input checked="" type="checkbox"/> Horizontal Committee | 01.10.2021 |
| | <input checked="" type="checkbox"/> EU PPE Expert Group | 18.11.2022 |
| Question related to <input checked="" type="checkbox"/> PPE Regulation <input type="checkbox"/> PPE Guidelines <input checked="" type="checkbox"/> EN/prEN: General <input type="checkbox"/> Other: | | |
| Article: | Annex: | Clause: |
| Key words: Antineoplastic agents | | |
| Question: Is it possible to use the phrase “against chemicals risks and antineoplastic agents” in the scope of EU type certificate according to EN ISO 374-1: 2016, when the glove is tested with permeation test, EN 374-3, using an appropriate number of antineoplastic agents? Could it be possible when the glove is tested against at least four of antineoplastic agents chosen from this list? Cyclophosphamide / Carmustine / Adryamicin (Doxorubicin or Adriblastine) / Fluorouracil / Methotrexate / Vincristine / cis Platinum / Daunorubicin. This list represents the most used chemicals in hospital treatments. | | |
| Solution: Yes. The phrase can be used if protection against an anti-neoplastic agent is demonstrated. The list of agents tested shall be included on the Certificate and the User Information. | | |