



**CO-ORDINATION OF NOTIFIED BODIES**  
**PPE Regulation 2016/425**

PPE-R/05.05-316  
Version 02

**RECOMMENDATION FOR USE**

Number of pages: 1	Approval stage:	Approved on:
Origin: Vertical Group 5	<input checked="" type="checkbox"/> Vertical Group	15.06.2021
	<input checked="" type="checkbox"/> Horizontal Committee	01.10.2021
	<input checked="" type="checkbox"/> EU PPE Expert Group	18.11.2022
Question related to <input checked="" type="checkbox"/> PPE Regulation <input type="checkbox"/> PPE Guidelines <input checked="" type="checkbox"/> EN/prEN: EN 366 / EN ISO 6942 <input type="checkbox"/> Other:		
Article:	Annex:	Clause:
Key words: Blackening of calorimeter		
Question: In EN 366 / EN ISO 6942 it is said that the calorimeter shall be blackened before the tests. Is this absolutely necessary? If the answer is YES, what type of paint?		
Solution: YES, it is necessary. In EN 367:1992 the following information is given: Black paint: Nextel Velvet Coating: Black 2010 3M UK Ltd. P.O. Box 38 Yeoman House 63, Croydon Road, Penge London SE 20 7TR United Kingdom Paint remover: Acetone		